



**INSTITUT ZA KARDIOVASKULARNE BOLESTI
VOJVODINE, SREMSKA KAMENICA**
Put doktora Goldmana 4, 21204 Sremska Kamenica
REPUBLIKA SRBIJA
Tel: 021/48 05 700; Fax: 021/66 22 881



JN-USL-03/17
Broj: 1548-4
Datum: 12.5.2017.
Sr. Kamenica

PREDMET: Dodatno pojašnjenje u vezi sa pripremanjem ponuda za javnu nabavku br. JN-USL-03/17 – Popravke i održavanje medicinske opreme

Dana 12.5.2017. godine elektronskim putem nam je dostavljen dopis zainteresovanog lica kojim se traži dodatno pojašnjenje u vezi sa pripremanjem ponude za javnu nabavku JN-USL-03/17 – Popravke i održavanje medicinske opreme, u okviru kojeg je postavljeno sledeće pitanje:

1. Proizvođač infuzionih i špric pumpi Arcomed Ag, Švajcarska, izvršio je promenu šifri rezervnih delova, tako da se nove, validne šifre ne slažu sa šiframa koje su objavljene u obrascu ponude konkursne dokumentacije. Takođe, predlažemo da dopunite obrazac ponude sa dodatnim rezervnim delovima koje je prema iskustvu proizvođača i ovlašćenog servisera Arcomed Ag, potrebno zameniti usled kvara i redovnog servisa.

Predlog izmene i dopune rezervnih delova sa odgovarajućim šiframa:

Medicinska oprema: Syramed uSP6000 Classic

71272 House maintenance Kit
70243 Ribbon cable complete
99245 Fusing kit
70331 Door cover kit
70612 Drive head assembly kit
70516 Clutch drive kit
70823 Housing front
70835 Housing rear complete
70911 Main PCB programmed
70300 Display PCB
71175 Back Plate with nut
70357 Door Lock Support
70260 Door external part complete
73350 Door internal part with magnet
73070 Gearwheel no. 3
73080 Gearwheel no. 5
98901 Rubber foot front blue
71231 Labels 8mm
71160 Battery NiMeH 9,6V/2150mAh
71210 Lithium battery time keeper

Medicinska oprema: Volumed uVP7000 Classic

75046 Air in line detector kit
76311 Door pressure kit PVC
75094 StopFlow Robson complete with adapter
75610 Main PCB V1.07 with plug holder
75497 Display PCB complete
75490 Display PCB V1.06 complete
75005 Housing front with EC-holder

75007 Housing front with EC-holder Upstream
98502 Easy Clip

Medicinska oprema: Volumed uVP5005

80920 Housing rear internal drop detector
80930 Housing front
80940 Pumpblock repair kit
98601 Pressure sensor PVC/silicone
99003 Main PCB complete V2.06
99020 Display PCB complete
94328 Door complete
98404 Front plate
99840 Circlips for shafts bulk 3.2
93380 Battery Ni-Cd 12V/1200mAh Ni-Cd

Na osnovu člana 63. stav 3. Zakona o javnim nabavkama („Službeni glasnik RS“ br. 124/2012, br. 14/15 i 68/15), dostavljamo odgovore na postavljena pitanja:

1. Naručilac će izvršiti izmenu konkursne dokumentacije.

Ova informacija, shodno članu 63. stav 3. Zakona o javnim nabavkama, se objavljuje na Portalu javnih nabavki i internet stranici Naručioca.

Komisija za javnu nabavku